

Suggestions for performance evaluation with the **LUCAS® 3** Chest Compression System

Name:		Unit:		
Reviewer:				
sys safe	s Performance Evaluation is a suggested basic assessment. This is a limited evaluation and does not cover a ely. It is designed to be completed after viewing the appropagualified instructor. Refer to the LUCAS 3 device Instructor.	ment of a train all the informa priate inservice	ee's ability to o ation and skills video or observi	required to operate this device ng an equipment demonstration given
Pe	rformance criteria	Complete	Incomplete	Comments
1.	Opens case; presses ON/OFF for 1 second to start self-test and power up the LUCAS 3 device.			
2.	Removes back plate from the case.			
3.	Verbalizes "stop manual CPR" to other rescuer (instructor may "act" as other rescuer).			
4.	Places back plate under patient, immediately below the patient's armpits.			
5.	Verbalizes "resume manual CPR" to other rescuer.			
6.	Lifts upper part of the LUCAS 3 device from case and pulls once on the release rings to check that the claw locks are open, then lets go of rings.			
CO	structor is prepared and starts timer when manual mpressions stop. Connects the upper part to the back plate, starting on the side closest to user. Listens for click.			
8.	Pulls up once to ensure attachment.			
9.	Positions the suction cup immediately above the end of the sternum in the center of the chest. Adjusts if necessary.			
10	Pushes the suction cup down using two fingers with the LUCAS 3 device in the ADJUST mode until the pressure pad inside suction cup touches the patient's chest without compressing the chest.			
11	. Pushes PAUSE to lock start position.			
12	Presses ACTIVE (30:2) or (continuous). Instructor stops timer: seconds (goal: ≤ 20 seconds).			
13	. Fully extends the stabilization strap at the buckles.			
14	. Places support cushion under patient's neck.			
15	Connects buckles on support cushion straps to device straps.			
16	. Tightens support cushion straps firmly.			
17	Visually checks for proper position of suction cup, adjusting if needed.			

General Performance Guidelines

Consider the following issues when determining competency:

- Was application and operation performed in a safe manner?
- Were compressions started within 20 seconds from the time manual compressions were stopped to connect the upper part of the LUCAS device?
- Able to verbalize contraindications and side effects?

Suggested Advanced Performance Scenarios

• Change battery during patient use

Optional Activities

- Apply defibrillation pads-performs defibrillation
- Perform ventilations with bag-valve-mask
- Perform ventilations with advanced airway

Evaluation and action plan					

Refer to operating instructions for complete directions for use indications, contraindications, warnings, cautions, and potential adverse events.

Physio-Control is now part of Stryker.

For further information, please contact Physio-Control at 800.442.1142 (U.S.), 800.668.8323 (Canada) or visit our website at www.physio-control.com

Physio-Control Headquarters 11811 Willows Road NE

11811 Willows Road NE Redmond, WA 98052 www.physio-control.com

Customer Support
P. O. Box 97006
Redmond, WA 98073
Toll free 800 442 1142
Fax 800 426 8049

Physio-Control Canada

Physio-Control Canada Sales, Ltd. 45 Innovation Drive Hamilton, ON L9H 7L8 Canada Toll free 800 668 8323 Fax 877 247 7925